



Analysis by Havard Industries

Certificate of Analysis

Sample ID: D1F6

Batch:

Date Tested: 8/23/2021



Sample pH: 7.15

Notes:

Samples were extracted into water with heat and shaking. Samples were allowed to cool back to room temperature before measuring. pH meter was calibrated using a 4-point calibration and produced a slope of 97%+. A control sample was used to measure the relative difference to house RO water. Control sample was measured 7.13.

Havard Industries
6300 Boucher Dr.
Edmond, OK
(405) 888-0961

Jeffery Havard
Lab Manager, Havard Industries

Lic# LAAA-8SPC-5FH4

This report must be reproduced in full except upon approval by the lab. NT - Not tested, NR - Not Run.